

AN APPARATUS AND METHOD FOR DETECTING RADIATION  
THAT USES A STIMULABLE PHOSPHOR

BACKGROUND OF THE INVENTION

Field of the invention

5           This invention relates to an apparatus and a method  
for radiation measurement that are suitable for use within  
and outside nuclear reactor or accelerator facilities when  
there is a need to perform real-time monitoring of  
radiation intensity over a wide dynamic range at many and  
10   broadly distributed sites of measurement. In nuclear  
reactor and accelerator facilities, one must use a  
detection system with which radiation doses ranging from a  
very weak level to a very intense level can be monitored at  
all times in various places. A portable radiation monitor  
15   is also required for personal radiation exposure  
management.

          To meet these needs, the present inventors developed  
a simple differential and integral-type apparatus and  
method for radiation measurement that were based on the  
20   combination of two principles, one performing constant  
radiation monitoring and the other integrating measured  
radiations with a stimuable phosphor. The present  
inventors also developed an apparatus and method for  
radiation measurement that was applicable to a remote  
25   sensing radiation measuring apparatus, a radiation  
distribution measuring apparatus and a portable radiation  
measuring apparatus.

Prior arts

Conventionally, radiation measurement has been performed with an ionization chamber, a Geiger-Mueller counter (GM counter) and a scintillation detector consisting of a scintillator combined with a photomultiplier tube whereas neutron dosimetry has been conducted with a  $\text{BF}_3$  or  $^3\text{He}$  counter. However, measuring a wide dynamic range of radiation doses from a very weak level to a very intense level by a single detector is extremely difficult and to overcome this difficulty, an ionization chamber of low sensitivity has been used in combination with a scintillation detector of high sensitivity. High-intensity radiation doses that occur instantaneously around an accelerator or a target or which result from an unexpected accident in nuclear reactor facilities have also been difficult to measure since they saturate most detectors.

A stimuable phosphor has two actions, one is accumulating an incident radiation and outputting the quantity of accumulated radiation as stimulated fluorescence upon stimulation by exciting light and the other is emitting prompt fluorescence upon excitation by an incident radiation. On the basis of these two actions, a method was developed that could measure the quantity of incident radiation by selectively detecting stimulated fluorescence and prompt fluorescence at specified time intervals [Japanese Patent Application No. 50301/1999]. By using an apparatus for radiation measurement that relies upon this technique of time-division multiplexing, a wide dynamic range of radiation doses from a very low level to a

very high level can be covered with a single detector and, at the same time, instantaneous high-intensity radiation and neutron doses can be measured.

One of the important constituents of the present invention is a stimuable phosphor used as a radiation detection medium. Fig. 35 is a schematic of the apparatus descried in D. J. Huntley et al., Nature, Vol. 313, 10, pp. 105-107 for reading the quantity of radiation accumulated in the stimuable phosphor. A stimuable phosphor sheet illuminated with a radiation in a different place is set on a reading table and then illuminated from the front with exciting light issued from an argon laser; the resulting emission of stimulated fluorescence is passed through bandpass optical filters centered at the frequency of stimulated fluorescence and subsequently detected with a photomultiplier tube; in accordance with its intensity, the detected fluorescence is digitized by a signal processor to determine the quantity of incident radiation.

In another conventional method, a small amount of stimuable phosphor is attached to the end of an optical fiber and illuminated with exciting light launched into the optical fiber and photons in the stimulated fluorescence that emits during the irradiation are counted to determine the incident radiation dose. This method was described by Kitaguchi et al. in JAERI-Conf 98-011, pp. 62-66 and the apparatus for implementing it is illustrated in Fig. 36. Due to the small amount of the stimuable phosphor used, it is difficult to increase the sensitivity of this method.

According to Radiation Protection Dosimetry, Vol. 65, No. 1/4, pp. 267-272, S. W. S. McKeever et al. proposed a method in which a stimuable phosphor was illuminated with exciting light from a pulsed light source to read the dose of accumulated radiation. A schematic of this pulsed OSL (optically stimulated luminescence) system is shown in Fig. 37. After applying an impulse of exciting light on the timing shown in Fig. 38, the stimulated luminescence emitted in accordance with the life of fluorescence is amplified with a high-speed amplifier and photons in it are counted with a fast counter circuit to determine the accumulated dose of radiation. This method requires a high-speed measurement system including a photodetector.

A conventional method that can measure the dose of radiation and the position of measurement simultaneously with a simple apparatus was described by Maekawa et al. in Hoshasen, Vol. 21, No. 3, pp. 69-78. In this method, the detecting portion shown in Figs. 39 and 40 is installed at each site of measurement and a plurality of such detecting portions are connected in series by an optical fiber for radiation measurement. Scintillator is used as a radiation detecting medium and the fluorescence emitted is detected with a wavelength shifting optical fiber and the radiation intensity as well as the position of measurement are determined from the difference between the arrival times of fluorescence at opposite ends of the wavelength Shifting optical fiber. A problem with this method is that if a radiation of high intensity is incident, not only its

intensity but also the position of measurement is difficult to determine.

#### SUMMARY OF THE INVENTION

An object, therefore, of the present invention is to  
5 provide an apparatus and a method for radiation measurement  
which allow a wide dynamic range of radiation doses from a  
very weak level to a very intense level to be covered with  
a single detector and which enable simple, sensitive and  
correct dosimetry of high-intensity radiations and neutrons  
10 that occur instantaneously around an accelerator or a  
target or which result from an unexpected accident in  
nuclear reactor facilities.

Another object of the invention is to provide a  
compact and simple apparatus for radiation measurement that  
15 can be used in routine operations of nuclear reactor and  
accelerator facilities, as well as a method for use on the  
apparatus.

Yet another object of the invention is to provide an  
apparatus and a method that enable simple and sensitive  
20 measurement of radiation doses at many and broadly  
distributed sites of measurement.

A further object of the invention is to provide an  
apparatus for radiation measurement with which a two-  
dimensional radiation image can be easily obtained.

#### 25 BRIEF DESCRIPTION OF THE DRAWINGS

Fig. 1 shows an apparatus for implementing the method  
of radiation measurement recited in claim 1 which relies  
upon time-division multiplexing of stimulated fluorescence

and prompt fluorescence that are emitted from a stimuable phosphor;

Fig. 2 is a timing chart for time-division multiplexing that is performed by switching between a stimulated fluorescence mode and a prompt fluorescence mode;

Fig. 3 shows an apparatus for implementing the method of radiation measurement recited in claim 2 which is capable of detecting high-intensity radiation with a stimuable phosphor;

Fig. 4 shows the differential and integral-type apparatus for radiation measurement recited in claim 3 which relies upon reading the light of saturation detecting LED;

Fig. 5 is a timing chart for reading stimulated fluorescence and making surveillance for system saturation on the basis of time-division multiplexing;

Fig. 6 shows the differential and integral-type apparatus for radiation measurement recited in claim 4 which relies upon reading stimulated fluorescence with an altered sensitivity for fluorescence detection;

Fig. 7 shows the differential and integral-type apparatus for radiation measurement recited in claim 5 which relies upon reading stimulated fluorescence with an altered quantity of exciting light;

Fig. 8 shows the differential and integral-type apparatus for radiation measurement recited in claim 6 which relies upon correcting the effect of prompt

fluorescence that is incident by chance during radiation measurement with stimulated fluorescence;

Fig. 9 shows an apparatus for implementing the method of radiation measurement recited in claim 7 which reads the dose of accumulated radiation in a stimuable phosphor upon illumination with pulsed exciting light;

Fig. 10 is a timing chart for signals generated in the method of radiation measurement recited in claim 7 which reads the dose of accumulated radiation in a stimuable phosphor upon illumination with pulsed exciting light;

Fig. 11 shows an apparatus for implementing the method of radiation measurement recited in claim 8 which reads the dose of accumulated radiation in a stimuable phosphor using a gated photomultiplier tube as a photodetector;

Fig. 12 shows an apparatus for implementing the method of radiation measurement recited in claim 9 which reads the dose of accumulated radiation in a stimuable phosphor by coincident counting of photons in pulsed exciting light and stimulated fluorescence signals;

Fig. 13 is a timing chart for coincident counting of stimulated fluorescence signals and reading signals constructed from pulsed exciting light;

Fig. 14 shows an apparatus for implementing the method of radiation measurement recited in claim 10 which uses a stimuable phosphor to be illuminated with exciting light through laterally radiating optical fibers;

Fig. 15 shows an apparatus for implementing the method of radiation measurement recited in claim 11 which uses a stimuable phosphor to be illuminated with exciting light through semi-laterally radiating optical fibers;

5           Fig. 16 shows an apparatus for implementing the method of radiation measurement recited in claim 12 which uses a stimuable phosphor to be illuminated with exciting light through semi-laterally radiating optical fibers in the presence of a light reflector at their back;

10           Fig. 17 shows a radiation detecting portion for use in the method of claim 13 which employs a stimuable phosphor in combination with a laterally radiating optical fiber and wavelength shifting optical fibers sensitive to the wavelength of fluorescence;

15           Fig. 18 shows a radiation detecting portion for use in the apparatus of claim 14 which is a planar array consisting, in superposition, of laterally radiating optical fibers, a stimuable phosphor, an optical filter and wavelength shifting optical fibers;

20           Fig. 19 shows an apparatus for radiation measurement that uses the radiation detecting portion shown in Fig. 18;

          Fig. 20 shows a radiation detecting portion for use in the apparatus of claim 15 which is a ribbon array of laterally radiating optical fibers sandwiched between  
25   assemblies each consisting of a stimuable phosphor, an optical filter and a ribbon array of wavelength shifting optical fibers;

          Fig. 21 shows a radiation detecting portion for use



in the apparatus of claim 16 which is a ribbon array of laterally radiating optical fibers sandwiched between assemblies consisting of different stimuable phosphors, an optical filter and a ribbon array of wavelength shifting  
5 optical fibers;

Fig. 22 shows a radiation detecting portion for use in the apparatus of claim 17 which is a cylindrical arrangement of a laterally radiating optical fiber, a stimuable phosphor, an optical filter and wavelength  
10 shifting optical fibers;

Fig. 23 shows the apparatus for radiation measurement recited in claim 18 which enables measurement of a positional distribution of radiation using a multiple of series connected radiation detecting portions including a  
15 stimuable phosphor;

Fig. 24 shows the apparatus for radiation measurement recited in claim 19 which enables measurement of a positional distribution of radiation using a multiple of radiation detecting portions including a stimuable  
20 phosphor that are series connected with optical delay mechanisms;

Fig. 25 shows the apparatus for radiation measurement recited in claim 20 which enables measurement of a positional distribution of radiation with pulsed exciting  
25 light using a radiation detecting portion in the form of a planar array;

Fig. 26 shows the apparatus for radiation measurement recited in claim 21 which enables measurement of a

positional distribution of radiation with pulsed exciting light using a radiation detecting portion in the form of a planar array and a streak camera;

Fig. 27 shows the apparatus for radiation measurement  
5 recited in claim 23 which enables measurement of a positional distribution of radiation with a single bundle of wavelength shifting optical fibers that read the stimulated fluorescence emitted from a multiple of stimuable phosphors provided at predetermined sites of  
10 radiation measurement;

Fig. 28 shows the apparatus for radiation measurement recited in claim 24 which produces a two-dimensional radiation image using laterally radiating optical fibers to launch exciting light;

15 Fig. 29 shows the apparatus for radiation measurement recited in claim 25 which uses the same photodetector to measure the fluorescence output from both ends of a wavelength shifting optical fiber bundle in a radiation detecting portion in the form of a planar array;

20 Fig. 30 shows the differential and integral-type apparatus for radiation measurement recited in claim 26 which uses a radiation detecting portion in the form of a planar array;

Fig. 31 shows a differential and integral-type  
25 apparatus for implementing the method of radiation measurement recited in claim 27 which applies a pulsed optically stimulated fluorescence reading technique to a radiation detecting portion in the form of a planar array;

Fig. 32 shows the differential and integral-type apparatus for radiation measurement recited in claim 28 which is capable of neutron measurement using both a neutron converter and a stimuable phosphor as a radiation  
5 detecting medium;

Fig. 33 shows an apparatus for implementing the method of radiation measurement recited in claim 29 which is capable of neutron measurement using a fast neutron moderator;

10 Fig. 34 shows the differential and integral-type apparatus for radiation measurement recited in claim 30 which is characterized by correcting the accumulated dose of radiation in a stimuable phosphor on the basis of its temperature as measured with a temperature sensor;

15 Fig. 35 is a schematic of an apparatus for implementing a conventional method of reading stimulated fluorescence;

Fig. 36 is a schematic of an apparatus for implementing another conventional method of radiation  
20 measurement using a radiation detecting portion having a small amount of stimuable phosphor attached to the end of an optical fiber;

Fig. 37 is a schematic of an apparatus for implementing yet another conventional method of reading the  
25 accumulated dose of radiation in a stimuable phosphor using a pulsed exciting light source;

Fig. 38 is a diagram showing the timing of reading stimulated fluorescence by illumination with pulsed

exciting light in a conventional method, as well as the emission characteristics of stimulated fluorescence obtained by a conventional fast photon counting method;

Fig. 39 is a schematic of an apparatus for  
5 implementing a conventional method of measuring the dose of radiation and the position of measurement simultaneously using a scintillator and an optical fiber; and

Fig. 40 is a schematic of a conventional radiation detecting portion capable of simultaneous measurement of  
10 the dose of radiation and the position of measurement using a scintillator and a wavelength shifting optical fiber.

#### DETAILED DESCRIPTION OF THE INVENTION

While making efforts to develop an apparatus for radiation measurement in which a wide dynamic range of  
15 radiation doses from a very low level to a very high level can be covered with a single detector and which can also measure high-intensity radiation and neutron doses, the present inventors noted two actions of a stimuable phosphor, one being the ability to accumulate an incident  
20 radiation and output the quantity of accumulated radiation as stimulated fluorescence upon stimulation by exciting light and the other being the ability to emit prompt fluorescence upon excitation by an incident radiation, and developed a method that could measure the quantity of  
25 incident radiation by selectively detecting stimulated fluorescence and prompt fluorescence at specified time intervals [Japanese Patent Application No. 50301/1999, supra]. By employing this method, the required radiation

detecting performance can be easily ensured by a simple apparatus. Another purpose of the invention is to provide an apparatus for radiation measurement in which if a fluorescence detecting mechanism is saturated to failure by the incidence of an intense radiation within a short period of time, stimulated fluorescence is read after the recovery of the fluorescence detecting mechanism so as to measure the incident dose of intense radiation. To this end, the fluorescence detecting mechanism is monitored for saturation on a time-division multiplexing basis and the dose of radiation accumulated in a stimuable phosphor is measured with the quantity of exciting light or the sensitivity for fluorescence detection being altered in such a way as to ensure that the fluorescence detecting mechanism is not saturated to failure when reading the rapid intense dose of incident radiation after recovery from the saturated state.

To read the dose of radiation accumulated in a stimuable phosphor, it is illuminated with short durations of pulsed exciting light and signal processing is performed with a conventional radiation detecting system by taking advantage of the short lifetime of the stimuable phosphor. This method offers the advantage that the accumulated dose of radiation can be conveniently measured by easily switching between stimulated fluorescence and prompt fluorescence on a time-division multiplexing basis.

If a laterally radiating optical fiber is used to illuminate the stimuable phosphor with exciting light,

convenient and remote-controlled excitation can be performed for reading the dose of radiation accumulated in the stimuable phosphor. This eliminates the need to install an exciting light source such as a laser in  
5 intensely irradiated areas and, what is more, the mechanism for illuminating exciting light can be constructed in a compact size. The use of a laterally radiating optical fiber has the added advantage of providing ease in measuring a two-dimensional radiation image using a  
10 stimuable phosphor sheet.

To ensure a large area for detecting both stimulated fluorescence and prompt fluorescence, wavelength shifting optical fibers sensitive to the wavelength of fluorescence may be arranged side by side and this eliminates the need  
15 to install a photodetector such as a photomultiplier tube in intensely irradiated areas. If desired, a ribbon array of wavelength shifting optical fibers may be used in combination with a parallel arrangement of laterally radiating optical fibers. The laterally radiating optical  
20 fibers may also be sandwiched between assemblies each consisting of a stimuable phosphor, an optical filter and a ribbon array of wavelength shifting optical fibers. These designs allow for constructing small and various shapes of radiation detecting portion.

25 Alternatively, the detecting portion of the sandwich structure may be extended over a distance and exciting light having very short pulse widths is launched into laterally radiating optical fibers so that a stimuable

phosphor is illuminated with the propagating short-pulsed exciting light to emit stimulated fluorescence which is then detected with wavelength shifting optical fibers.

This method enables measurement of a long-distance

5 distribution of radiation dose. If this method and detection of prompt fluorescence are selectively used, the distribution of radiation dose can be measured in many and broadly distributed areas.

If desired, the stimuable phosphor used as a  
10 radiation detecting medium in the methods described above may incorporate at least one neutron converter selected from among Gd,  $^6\text{Li}$  and  $^{10}\text{B}$  as materials that convert neutrons to an ionizable radiation. Alternatively, the stimuable phosphor may be mixed or combined with such  
15 neutron detecting medium. In either way, one can provide an apparatus and a method for radiation measurement that are capable of neutron detection.

#### Examples

Claim 1 sets forth the basic concept of the present  
20 invention and an example of it is described below with reference to Fig. 1. As already mentioned, a stimuable phosphor which is used as a radiation detecting medium in the invention has two actions, one is accumulating an incident radiation and outputting the quantity of  
25 accumulated radiation as "stimulated fluorescence" upon stimulation by exciting light and the other is emitting "prompt fluorescence" upon excitation by an incident radiation. The basic concept of the invention is to

provide a method of measuring the quantity of incident radiation by selectively detecting stimulated fluorescence and prompt fluorescence real-time at specified time intervals using a fluorescence detecting mechanism.

5           In the example under consideration, a sheet of radiation detecting medium can be used, as exemplified by X-ray imaging plates of BAS Series commercially available from Fuji Photo Film Co., Ltd. The rear faces of the commercial X-ray imaging plates are not transparent but can  
10 be easily modified to become transparent. The stimuable phosphor used in these imaging plates is  $\text{BaFBr:Eu}^{2+}$  and can be excited in a wavelength band of 490 nm - 680 nm (where more than half of the maximum efficiency is exhibited). Upon illumination with exciting light,  $\text{BaFBr:Eu}^{2+}$  emits  
15 stimulated fluorescence at a wavelength of 390 nm and the life of emitted fluorescence is 0.8  $\mu\text{s}$ .

While the example under consideration assumes the use of  $\text{BaFBr:Eu}^{2+}$ , other stimuable phosphors such as  $\text{KCl:Eu}^{2+}$ ,  $\text{RbBr:Tl}$ ,  $\text{SrS:Eu}$  and  $\text{Sm}$  may be used and imaging plates  
20 containing these as detection mediums can be used as specimen by altering the wavelength of exciting light and the wavelength of stimulated fluorescence to be read.

A laser is used as the light source of exciting light. The required laser power depends on the reading  
25 speed and the reading area of the imaging plate in the form of a stimuable phosphor sheet. If the detection area is a circle with a diameter of 5 mm, a power of at least several milliwatts is required. Since  $\text{BaFBr:Eu}^{2+}$  in the imaging



plates of BAS Series (Fuji Photo Film Co., Ltd.) can be excited in a wavelength band of 490 nm - 680 nm, a laser operating in this wavelength band to produce high power with ease is either a semiconductor laser (635 nm) or a  
5 green-light laser (532 nm).

An optical bandpass filter centered at the wavelength of stimulated fluorescence is provided at the back of the imaging plate and the stimulated fluorescence emitted from the rear face of the imaging plate is passed through the  
10 filter to reject the unwanted scattering component of the exciting light. In the example under consideration, the stimulated fluorescence has a wavelength of 390 nm, so the optical filter should have a central wavelength of 390 nm.

We now describe the fluorescence detecting mechanism.  
15 The photodetector that can be used in the invention is a photomultiplier tube having high sensitivity around 390 nm which is the wavelength of stimulated fluorescence. The optical signal detected by the photomultiplier tube is fed into a signal processing circuit where it is amplified and  
20 subjected to other signal processing. In the example under consideration, the optical signal is processed by photon counting. Specifically, the optical signal is subjected to the shaping of pulse waveform, discriminated from noise in a pulse height discriminator and picked up as a pulse  
25 signal. Alternatively, the reading method conventionally applied to imaging plates may be employed such that a train of stimulated fluorescence signals are integrated in an integrating circuit to read the accumulated dose of

radiation.

The output pulse signal from the signal processing circuit is inputted to a mode switching circuit which, in response to a mode switching signal from a control circuit, supplies the pulse signal to either a stimulated fluorescence data collecting circuit if the operation is in a stimulated fluorescence mode, or a prompt fluorescence data collecting circuit if the operation is in a prompt fluorescence mode. In the example under consideration, the two data collecting circuits are each composed of a counting circuit.

Switching between the stimulated and the prompt fluorescence mode is performed by the control circuit on the timing shown in Fig. 2. While the imaging plate in the form of a stimuable phosphor sheet is illuminated with exciting light, stimulated fluorescence is read but, otherwise, prompt fluorescence is read. Hence, by controlling the exciting light source to emit exciting light synchronously with the reading cycles, both stimulated fluorescence and prompt fluorescence can be detected with a single fluorescence detecting mechanism at specified time intervals.

We now describe an example of claim 2 with reference to Fig. 3. When stimulated fluorescence and prompt fluorescence are being detected from the imaging plate by the fluorescence detecting mechanism at specified time intervals during radiation measurement in or around an accelerator, an intense radiation may be incident within a

short time to saturate the fluorescence detecting mechanism which is no longer operational. If this occurs, no signal enters the signal processing circuit. Hence, monitoring is performed with a saturation detecting circuit which, when  
5 the fluorescence detecting mechanism is saturated, sends a saturation signal to the control circuit. When the incidence of the rapid and intense radiation has ended and the fluorescence detecting mechanism has recovered from saturation, the operation is switched to the stimulated  
10 fluorescence mode by the mode switching circuit and the stimulated fluorescence is detected, thereby measuring the rapid and intense dose of the incident radiation.

We now describe an example of claim 3 with reference to Fig. 4. The apparatus for radiation measurement shown  
15 in Fig. 4 comprises the following components: a stimuable phosphor serving as a radiation detecting medium; an exciting light source for illuminating the stimuable phosphor with exciting light to read the accumulated dose of radiation at specified time intervals; an optical  
20 bandpass filter centered at the wavelength of fluorescence; a photodetector for detecting stimulated fluorescence and prompt fluorescence; a signal processing circuit for amplifying and processing a stimulated fluorescence signal or a prompt fluorescence signal that are output from the  
25 photodetector; a mode switching circuit for switching between the stimulated fluorescence signal and the prompt fluorescence signal at specified time intervals; a saturation detecting circuit for monitoring the operation

of the fluorescence detecting mechanism; an LED for use in saturation detection; a circuit for altering the quantity of exciting light; a control circuit for controlling the exciting light source, the mode switching circuit, the  
5 saturation detecting circuit and the light quantity altering circuit at specified time intervals; a stimulated fluorescence data collecting circuit for measuring stimulated fluorescence; a prompt fluorescence data collecting circuit for measuring prompt fluorescence; and  
10 a data processing circuit which collects stimulated and prompt fluorescence data at specified time intervals synchronously with the control circuit and determines the dose of incident radiation on the basis of the collected data.

15       The following are details of this apparatus for radiation measurement. In the example under consideration, a sheet of radiation detecting medium can be used, as exemplified by X-ray imaging plates of BAS Series commercially available from Fuji Photo Film Co., Ltd.  
20 The rear faces of the commercial X-ray imaging plates are not transparent but can be easily modified to become transparent. The stimuable phosphor used in these imaging plates is BaFBr:Eu<sup>2+</sup> and can be excited in a wavelength band of 490 nm - 680 nm (where more than half of the  
25 maximum efficiency is exhibited). Upon illumination with exciting light, BaFBr:Eu<sup>2+</sup> emits stimulated fluorescence at a wavelength of 390 nm. While the example under consideration assumes the use of BaFBr:Eu<sup>2+</sup>, other

stimulable phosphors such as  $\text{KCl:Eu}^{2+}$ ,  $\text{RbBr:Tl}$ ,  $\text{SrS:Eu}$  and  $\text{Sm}$  may be used and imaging plates containing these as detection mediums can be used as specimen by altering the wavelength of exciting light and the wavelength of  
5 stimulated fluorescence to be read.

A laser is used as the light source of exciting light. The required laser power depends on the reading speed and the reading area of the imaging plate in the form of a stimulable phosphor sheet. If the detection area is a  
10 circle with a diameter of 5 mm, a power of at least several milliwatts is required. Since  $\text{BaFBr:Eu}^{2+}$  in the imaging plates of BAS Series (Fuji Photo Film Co., Ltd.) can be excited in a wavelength band of 490 nm - 680 nm, a laser operating in this wavelength band to produce high power  
15 with ease is either a semiconductor laser (635 nm) or a green-light laser (532 nm). The quantity of exciting light to be issued from the laser can be altered by a control signal output from the light quantity altering circuit. The exciting light can also be controlled to turn on and  
20 off by an ON/OFF control signal from the control circuit.

An optical bandpass filter centered at the wavelength of stimulated fluorescence is provided at the back of the imaging plate and the stimulated fluorescence emitted from the rear face of the imaging plate is passed through the  
25 filter to reject the unwanted scattering component of the exciting light. In the example under consideration, the stimulated fluorescence has a wavelength of 390 nm, so the optical filter should have a central wavelength of 390 nm.

We now describe the fluorescence detecting mechanism. The photodetector that can be used in the invention is a photomultiplier tube having high sensitivity around 390 nm which is the wavelength of stimulated fluorescence.

5 Although not shown, a high bias voltage is applied to the photomultiplier tube. The light amplification factor depends on the applied bias voltage.

The optical signal detected by the photomultiplier tube is fed into the signal processing circuit where it is  
10 amplified and subjected to other signal processing. In the example under consideration, the optical signal is processed by photon counting. Specifically, the optical signal is subjected to the shaping of pulse waveform, discriminated from noise in a pulse height discriminator  
15 and picked up as a pulse signal. Alternatively, the reading method conventionally applied to imaging plates may be employed such that a train of stimulated fluorescence signals are integrated in an integrating circuit to read the accumulated dose of radiation.

20 The output pulse signal from the signal processing circuit is inputted to the mode switching circuit which, in response to a mode switching signal from the control circuit, supplies the pulse signal to either the stimulated fluorescence data collecting circuit if the operation is in  
25 a stimulated fluorescence mode, or the prompt fluorescence data collecting circuit if the operation is in a prompt fluorescence mode. In the example under consideration, the two data collecting circuits are each composed of a

counting circuit.

Switching between the stimulated and the prompt fluorescence mode is performed by the control circuit on the timing shown in Fig. 5. While the imaging plate in the form of a stimuable phosphor sheet is illuminated with exciting light, stimulated fluorescence is read but, otherwise, prompt fluorescence is read. Hence, by controlling the exciting light source to emit exciting light synchronously with the reading cycles, both stimulated fluorescence and prompt fluorescence can be detected with a single fluorescence detecting mechanism at specified time intervals.

When stimulated fluorescence and prompt fluorescence are being detected from the imaging plate by the fluorescence detecting mechanism at specified time intervals during radiation measurement in or around an accelerator, an intense radiation may be incident within a short time to saturate the fluorescence detecting mechanism which is no longer operational. If this occurs, the photodetector is saturated and fails to operate, so no signal enters the signal processing circuit.

To catch this phenomenon positively, an LED is used that emits light of a wavelength that passes through the optical filter and the pulsed light from the LED is launched into the photodetector on the timing shown in Fig. 5, with the output signal from the signal processing circuit being constantly monitored by the saturation detecting circuit to check to see if the photodetector can

detect the incident pulsed light. If not, the saturation  
detecting circuit finds the photodetector saturated and  
sends a saturation signal to the control circuit. The  
saturation detecting circuit continues the monitoring until  
5 the incidence of the rapid intense radiation has ended and  
the fluorescence detecting mechanism has recovered from  
saturation. When the saturated mechanism has restored to  
the normal state, the control circuit causes the mode  
switching circuit to send a mode switching signal,  
10 whereupon the operation is switched to the stimulated  
fluorescence mode and the photons in the stimulated  
fluorescence are counted to measure the dose of incidence  
of rapid and intense radiation.

When measuring the stimulated phosphor by this  
15 method, it is not known how much of the rapid and intense  
radiation has accumulated in the stimuable phosphor. To  
deal with this situation, the imaging plate is first  
illuminated with very weak exciting light and after knowing  
its quantity, the accumulated dose of radiation in the  
20 stimuable phosphor is read, thereby ensuring that the  
total dose of radiation accumulated in the stimuable  
phosphor is measured without causing saturation of the  
fluorescence detecting mechanism. Thus, according to claim  
3, the quantity of exciting light being applied to the  
25 stimuable phosphor is altered with the light quantity  
altering circuit on the basis of a signal from the control  
circuit while using the saturation detecting circuit as an  
auxiliary and this enables the dose of accumulated



radiation to be read from the stimuable phosphor without saturating the fluorescence detecting mechanism.

We now describe an example of claim 4 with reference to Fig. 6. In the previous example, the dose of an intense radiation that was incident within a short time is measured by counting the photons in stimulated fluorescence after fluorescence detecting mechanism has recovered from saturation. The same object (reading without saturation of the fluorescence detecting mechanism) can be attained by altering the sensitivity of the fluorescence detecting mechanism. In the example under consideration which uses a photomultiplier tube as the photodetector, a bias voltage altering circuit is used to alter the sensitivity for fluorescence detection and the intended operation is performed by controlling the bias voltage with the control circuit.

We now describe an example of claim 5 with reference to Fig. 7. This is the combination of claims 3 and 4 and characterized in that if an extremely large quantity of radiation is incident, the sensitivity for fluorescence detection is lowered by the sensitivity altering circuit and the time taken by measurement of stimulated fluorescence is shortened. According to claim 5, this and other methods may be used to optimize the reading of stimulated fluorescence.

We now describe an example of claim 6 with reference to Fig. 8. When the dose of incident radiation is measured at specified time intervals by counting the photons in

stimulated fluorescence in the foregoing examples using a mechanism for detecting stimulated fluorescence and prompt fluorescence, the imaging plate may incidentally be illuminated with a radiation during the reading process and counted in the result of measurement. To deal with this situation, the contribution from the prompt fluorescence is corrected with the data processing circuit on the basis of the dose of radiation measured at specified time intervals by counting the photons in prompt fluorescence using the fluorescence detecting mechanism and, as a result, the total dose of incident radiation can be correctly measured.

We now describe an example of claim 7 with reference to Fig. 9. The invention recited in claim 7 relates to a method applicable to an apparatus for radiation measurement using a stimuable phosphor and which is characterized by an improvement in the manner in which the dose of radiation accumulated in the stimuable phosphor is read.

The invention is particularly applicable to the case where a stimuable phosphor of short fluorescence lifetime ( $\leq 2 \mu\text{s}$ ) is used as a radiation detecting medium and characterized by applying the following signal processing technique which is conventionally used in processing signals from a semiconductor detector. In the example under consideration,  $\text{BaFBr:Eu}^{2+}$  with a fluorescence lifetime of  $0.8 \mu\text{s}$  is used as a radiation detecting medium.

To read the dose of radiation accumulated in this stimuable phosphor, the latter is illuminated with pulsed exciting light from a light source having an irradiation

time duration not longer than the lifetime of stimulated fluorescence. In the example under consideration, a pulsed green-light laser is used as a source of pulsed exciting light. This laser issues pulses of a very short duration  
5 of 2 ns. When the stimuable phosphor is illuminated with such pulsed exciting light, there occurs random emission of stimulated fluorescence as shown by timing diagram in Fig. 10. Since this emission practically ends within the lifetime of stimulated fluorescence, the emitted  
10 fluorescence is detected with a photodetector such as a photomultiplier tube and the detected signal is amplified with a charge sensitive preamplifier to produce an output signal also shown in the timing diagram. The output signal is inputted to a pulse shaping amplifier typically composed  
15 of an integrating and differentiating circuit so that it is subjected to both waveform shaping with a time constant longer than the lifetime of stimulated fluorescence from the stimuable phosphor and amplification. This procedure produces a pulse height signal which is an integral of the  
20 stimulated fluorescence emitted upon illumination with pulsed exciting light. The pulse height signal is then inputted to an analog/digital converter to determine the pulse height. The dose of radiation accumulated in the stimuable phosphor can be determined by a data collecting  
25 and processing circuit which integrates the pulse height obtained for each pulse of the exciting light that issues from the pulsed laser light source in response to a laser control signal sent from a pulse generator.

We now describe an example of claim 8 with reference to Fig. 11. The example is a modification of the radiation detecting apparatus recited in claim 3 and characterized by using a gated photomultiplier tube as the photodetector.

- 5 An example of the gated photomultiplier tube that can be used is R5916 manufactured by Hamamatsu Photonics. If BaFBr:Eu<sup>2+</sup> is used as a stimulable phosphor, the already mentioned green-light laser emitting pulses of a duration of 2 ns can be used as an exciting light source. In the
- 10 example under consideration, a gate signal having a time duration of 5 ns is input to the gated photomultiplier tube synchronously with the illumination of the stimulable phosphor with the pulsed exciting light such that the photomultiplier tube remains off as long as the
- 15 illumination continues but turns on when the gate is controlled after illumination with the exciting light. This procedure enables the stimulated fluorescence from the excited stimulable phosphor to be detected without being affected by illumination with the pulsed exciting light.
- 20 In other words, the radiation measuring apparatus recited in claim 8 is capable of determining the dose of radiation in a stimulable phosphor without being affected by illumination with the pulsed exciting light.

- We now describe an example of claim 9 with reference
- 25 to Fig. 12. Like claims 7 and 8, the invention recited in claim 9 relates to a method applicable to an apparatus for radiation measurement using a stimulable phosphor and which is characterized by an improvement in the manner in which

the dose of radiation accumulated in the stimuable phosphor is read. The improved method is particularly applicable to the case where a stimuable phosphor of short fluorescence lifetime ( $\leq 2 \mu s$ ) is used as a radiation  
5 detecting medium, provided that a moderate intensity of radiation has been accumulated in the stimuable phosphor. In the example under consideration,  $BaFBr:Eu^{2+}$  having a fluorescence lifetime of  $0.8 \mu s$  is used as the radiation detecting medium.

10 To read the dose of radiation accumulated in this stimuable phosphor, the latter is illuminated with pulsed exciting light from a light source having an irradiation time duration not longer than the lifetime of stimulated fluorescence. In the example under consideration, in order  
15 to read the dose of radiation accumulated in the stimuable phosphor having a fluorescence lifetime of  $0.8 \mu s$ , the stimuable phosphor is illuminated with pulsed exiting light from a light source capable of illumination for a time duration twice the lifetime of the fluorescence  
20 stimulated from the stimuable phosphor ( $0.8 \mu s \times 2 = 1.6 \mu s$ ). If the dose of the accumulated radiation is moderate, the timing of the emission of stimulated fluorescence is random and a single pulse of it may or may not be emitted synchronously with the pulsed exciting light as shown by  
25 timing diagram in Fig. 13. The emitted fluorescence is detected with a photodetector such as a photomultiplier tube and the detected signal is amplified with a preamplifier and a signal amplifier, from which an

amplified signal is output as also shown in the timing diagram. The output signal is inputted to a pulse height discriminator which outputs a digital signal also shown in the timing diagram. The output signal from the pulse  
5 height discriminator is fed into a coincident counting circuit and subjected to coincident counting with a read signal constructed by a read signal generator circuit using a signal indicating the time duration of illumination with the pulsed exciting light. This signal is constructed by a  
10 pulse generator as it generates a signal for controlling the laser from the pulsed exciting light source. The output signal from the coincident counting circuit corresponds to a single pulse of stimulated fluorescence that is being captured in response to illumination with the  
15 exciting light. The dose of radiation accumulated in the stimuable phosphor can be determined by counting the number of output signals from the coincident counting circuit with a counter circuit and subjecting the result to data processing in a data collecting and processing  
20 circuit.

We now describe an example of claim 10 with reference to Fig. 14. This is an improvement of a mechanism for illuminating the stimuable phosphor as a radiation detecting medium with exciting light to read the dose of  
25 radiation accumulated in the stimuable phosphor and is characterized by using a laterally radiating optical fiber as a radiator of exciting light in the mechanism. The example under consideration is the same as the method

recited in claim 7 which measures the dose of radiation accumulated in the stimuable phosphor, except that the exciting light from the pulsed laser light source is not directly applied to the stimuable phosphor but applied via the illustrated laterally radiating optical fiber. An example of the laterally radiating optical fiber that can be used is LUMINAS V Grade which is a laterally leaky optical fiber manufactured by Asahi Chemical Industry. A conventional optical fiber can be used as a laterally radiating type if a thin layer is removed from the entire circumference. The use of the laterally radiating optical fiber permits remote sensing by eliminating the need to install the exciting light source at the site of radiation detection.

We now describe an example of claim 11 with reference to Fig. 15. This example is a modification of claim 10, in which an optical fiber that radiates light from a portion of its circumference as shown enlarged in the inset is used as the lateral radiator in the mechanism for illuminating the stimuable phosphor with exciting light. A conventional optical fiber can be used as such semi-laterally radiating type if a thin layer is removed from a limited part of the circumference. The use of a semi-laterally radiating optical fiber eliminates the need to install the exciting light source at the site of radiation detection and permits efficient reading of the dose of radiation in the stimuable phosphor by radiating light only in a specified direction.

We now describe an example of claim 12 with reference to Fig. 16. This example is a modification of claim 11, in which the semi-laterally radiating optical fiber is overlaid with a light reflector which is remote from the light radiating part of the fiber so that leaking light is sufficiently reflected back to the radiating part to radiate as much light as possible. This design permits more efficient reading of the dose of radiation accumulated in the stimuable phosphor.

We now describe an example of claim 13 with reference to Fig. 17. The radiation detecting portion in this example uses a cylinder of stimuable phosphor; more specifically, it comprises in superposition an optical fiber capable of lateral radiation of light, a stimuable phosphor as a radiation detecting medium that is spaced from the optical fiber, an optical bandpass filter centered at the wavelength of fluorescence which is disposed outside the stimuable phosphor, and a multiple of wavelength shifting fibers that are sensitive to the wavelength of fluorescence and which are disposed around the optical filter to detect both stimulated fluorescence and prompt fluorescence. The outermost part of the radiation detecting portion is covered with a light shield. The stimuable phosphor is  $\text{BaFBr:Eu}^{2+}$  that emits stimulated fluorescence at 390 nm which has a lifetime of 0.8  $\mu\text{s}$ . In view of the wavelength of the stimulated fluorescence which is 390 nm in the example under consideration, a fluorescent plastic fiber having an excitation wavelength range of



320 nm - 395 nm and emitting fluorescence centered at the wavelength of 450 nm is used as the wavelength shifter. The wavelength shifted fluorescence has a lifetime not longer than 10 ns. The radiation detecting portion using the wavelength shifting fiber sensitive to stimulated fluorescence at 390 nm not only enables complete remote sensing but also increases the detecting area of the stimuable phosphor as a radiation detecting medium by a sufficient amount to enhance the sensitivity for radiation detection. The example under consideration can be used as a highly sensitive detector of a radioactive gas that is introduced into the space between the laterally radiating optical fiber and the stimuable phosphor.

We now describe an example of claim 14 with reference to Fig. 18. In this example, too, BaFBr:Eu<sup>2+</sup> having a fluorescence lifetime of 0.8  $\mu$ s is used as a stimuable phosphor. The radiation detecting portion in this example comprises the following components in planar superposition: four optical fibers with a diameter of 1 mm that are capable of lateral radiation of light; a stimuable phosphor as a radiation detecting medium; an optical bandpass filter centered at the wavelength of fluorescence; and four wavelength shifting fibers with a diameter of 1 mm that are sensitive to the wavelengths of stimulated fluorescence and prompt fluorescence and which are used to detect both types of fluorescence. An exemplary apparatus for measuring radiation using this radiation detecting portion is shown in Fig. 19. Using this radiation

detecting portion, one can perform complete remote sensing. In addition, the cross-sectional size of the radiation detecting portion can be reduced to less than 1 cm x 1 cm, which is small enough to permit easy radiation measurement even if the installation area is small or the detecting portion is quite long.

We now describe an example of claim 15 with reference to Fig. 20. The radiation detecting portion used in this example has a ribbon array of four optical fibers with a diameter of 1 mm that are capable of lateral radiation of light and which are sandwiched between two units of the following components in planar superposition: a stimuable phosphor as a radiation detecting medium; an optical bandpass filter centered at the wavelength of fluorescence; and four wavelength shifting fibers with a diameter of 1 mm that are sensitive to the wavelengths of stimulated fluorescence and prompt fluorescence and which are used to detect both types of fluorescence. Since the stimuable phosphor is provided on each side of the radiation detecting portion, not only can detection sensitivity be increased but also the direction of incidence of radiation can be estimated.

We now describe an example of claim 16 with reference to Fig. 21. As in the previous example, the radiation detecting portion has a ribbon array of four optical fibers with a diameter of 1 mm that are capable of lateral radiation of light and which are sandwiched between two units of the following components in planar superposition:

a stimuable phosphor as a radiation detecting medium; an optical bandpass filter centered at the wavelength of fluorescence; and four wavelength shifting fibers with a diameter of 1 mm that are sensitive to the wavelengths of stimulated fluorescence and prompt fluorescence and which are used to detect both types of fluorescence. The stimuable phosphor on one side is just the same as described above but the stimuable phosphor on the other side is adapted to be capable of detecting neutrons by mixing with Gd which is a neutron converter that converts neutrons to an ionizable radiation. This design allows for detection of neutrons in addition to ionizing radiations such as X-rays and gamma-rays. Some X-rays and gamma-rays are sensitive to the mixture of the neutron converter and the stimuable phosphor but their effects can be corrected on the basis of the data captured with the stimuable phosphor alone; hence, the dose of neutrons can be measured with high precision.

We now describe an example of claim 17 with reference to Fig. 22. In this example, too, BaFBr:Eu<sup>2+</sup> having a fluorescence lifetime of 0.8  $\mu$ s is used as a stimuable phosphor. The radiation detecting portion in this example is in a cylindrical form and comprises in superposition a single optical fiber with a diameter of 1 mm that is capable of lateral radiation of light, a stimuable phosphor as a radiation detecting medium, an optical bandpass filter centered at the wavelength of fluorescence, and a multiple of wavelength shifting fibers that are

sensitive to the wavelength of fluorescence and which are disposed around the optical filter to detect both stimulated fluorescence and prompt fluorescence. Using this radiation detecting portion, one can perform complete remote sensing. In addition, the cross-sectional size of the radiation detecting portion can be reduced to less than 1 cm in diameter, which is small enough to permit easy radiation measurement even if the installation area is small or the detecting portion is quite long.

We now describe an example of claim 18 with reference to Fig. 23. In this example, a stimuable phosphor spanning a long distance is used as a radiation detecting medium and in order to ensure that the dose of radiation accumulated in the stimuable phosphor can be read together with the associated position information, a special mechanism is used to illuminate the stimuable phosphor with exciting light of a very short pulse duration.

In order to enhance the position resolution for each site of radiation measurement in the example under consideration, the stimuable phosphor to be used must have a short enough fluorescence lifetime. This need can be met by  $\text{Y}_2\text{SiO}_5\text{:Ce}$  which emits stimulated fluorescence having a very short lifetime of 30 ns. The excitation wavelength of this stimuable phosphor is centered at 620 nm. Upon illumination with exciting light, this phosphor emits stimulated fluorescence at 410 nm. Since these characteristics are generally the same as those of  $\text{BaFBr:Eu}^{2+}$  which are used in the foregoing examples, the

exciting light source and the wavelength shifting optical fiber or fibers that are to be used in the example under consideration are also the same as what are used in the foregoing examples. A radiation detecting portion using the stimuable phosphor is provided at each site of measurement and the stimuable phosphor is illuminated with exciting light having a pulse duration no longer than 2 ns. The exciting light source is typically a semiconductor laser. The radiator of exciting light in the illumination mechanism is an optical fiber that has at least two areas spaced along its length where light can be radiated from the entire circumference or a part of its circumference. In registry with each site of light irradiation, the stimuable phosphor as a radiation detecting medium, an optical bandpass filter centered at the wavelength of fluorescence, and one or more wavelength shifting fibers that are sensitive to the wavelengths of stimulated fluorescence and prompt fluorescence and which are used to detect both types of fluorescence are superposed one on another to construct the radiation detecting portion. The time of incidence of pulsed exciting light output from the laterally radiating optical fiber is related to the temporal distribution of the stimulated fluorescence output from the wavelength shifting fibers and this relationship is used to determine the distribution of the dose of incident radiation at the specified sites of radiation measurement.

We now describe an example of claim 19 with reference

to Fig. 24. The radiation detecting portion of this example can be made of essentially the same components as in the example of claim 18. The radiator of exciting light in the mechanism for illuminating a stimuable phosphor (radiation detecting medium) with exciting light to read the dose of radiation accumulated in the stimuable phosphor is at least one illuminating optical fiber which consists of a laterally or semi-laterally radiating optical fiber, an optical delay mechanism and an ordinary optical fiber that are connected alternately. The back side of the laterally or semi-laterally radiating optical fiber length is overlaid with the stimuable phosphor as the radiation detecting medium and an optical filter centered at the wavelength of fluorescence which, in turn, is overlaid on the back side with at least one wavelength shifting optical fiber sensitive to stimulated fluorescence. The optical delay mechanism used in the example under consideration is an optical fiber having a length corresponding to a delay time of 30 ns. By increasing the interval between emissions of exciting light from successive laterally or semi-laterally radiating optical fiber lengths, the arrival time of exciting light is sufficiently delayed to ensure that the dose of radiation is correctly determined in each position of measurement.

We now describe an example of claim 20 with reference to Fig. 25. The radiation detecting portion used in this example is the same as what is used in the example of claim 14. In order to enhance the position resolution for each

site of radiation measurement, the stimuable phosphor to be used in the example under consideration must also have a short enough fluorescence lifetime. This need can be met by  $\text{Y}_2\text{SiO}_5\text{:Ce}$  which emits stimulated fluorescence having a very short lifetime of 30 ns. The excitation wavelength of this stimuable phosphor is centered at 620 nm. Upon illumination with exciting light, this phosphor emits stimulated fluorescence at 410 nm. Since these characteristics are generally the same as those of  $\text{BaFBr:Eu}^{2+}$  which are used in the foregoing examples, the exciting light source and the wavelength shifting optical fiber or fibers that are to be used in the example under consideration are also the same as what are used in the foregoing examples. The radiation detecting portion using the stimuable phosphor is provided at each site of measurement and the stimuable phosphor is illuminated with exciting light having a pulse duration no longer than 2 ns via at least one laterally radiating optical fiber. The exciting light source is typically a semiconductor laser. The stimulated fluorescence emitted from the stimuable phosphor is detected with the wavelength shifting fiber radiator. As in the example of claim 18, the time of incidence of pulsed exciting light output from the laterally or semi-laterally radiating optical fiber is related to the temporal distribution of the intensity of the stimulated fluorescence output from the wavelength shifting fibers and this relationship is used to determine the continuous distribution of the dose of incident

radiation at the specified sites of radiation measurement. To achieve this, the temporal distribution of the fluorescence output from the wavelength shifting fiber is measured with a signal processing circuitry which consists of a photomultiplier tube, a high-speed DC signal amplifier, an analog/digital converter, a memory circuit and a controlling/data collecting unit as shown in Fig. 25. The analog/digital converter should have a sampling rate of at least 100 MHz. The positional resolution that can be realized is determined by the lifetime of the fluorescence from the stimuable phosphor which is 30 ns in the example under consideration and the best value is about 10 m.

We now describe an example of claim 21 with reference to Fig. 26. This example is the same as the example of claim 20 (see Fig. 25) except that the signal processing circuitry for measuring the fluorescence output from the wavelength shifting fiber is replaced by a streak camera based technology. To be more specific, the fluorescence output from the wavelength shifting fiber is launched into a streak camera and the temporal distribution of the intensity of stimulated fluorescence is measured at high speed with varying time scan signals being supplied to the deflecting electrodes synchronously with the launching of pulsed exciting light into the semi-laterally radiating optical fiber. The measured temporal distribution is recorded in an imaging device such as a CCD camera. An example of the streak camera that can be used in the example under consideration is C2830 of Hamamatsu Photonics



which is capable of high-speed scanning. The recorded data is captured by the data collecting/controlling unit and the distribution of incident radiation doses at the sites of measurement can be determined on the basis of the temporal  
5 distribution of the captured fluorescence intensity data. Again, the positional resolution that can be realized is determined by the lifetime of the fluorescence from the stimuable phosphor which is 30 ns in the example under consideration and the best value is about 10 m.

10 We now describe the invention recited in claim 22. The apparatus described in connection with the examples of claims 18 - 21 can complete the intended measurement within a very short period of time. In the invention of claim 22, the procedure of such measurement is repeated more than  
15 once in order to read the signal for the radiation accumulated within the stimuable phosphor in the radiation detecting portion. If the radiation detecting portion is 100 m long, the reading process ends within 1  $\mu$ s. This means that even if the stimuable phosphor is illuminated  
20 with 1,000 applications of pulsed exciting light for reading the dose of accumulated radiation, the desired data can be collected within 1 ms. The number of times the pulsed exciting light is applied to the stimuable phosphor can be determined in accordance with the optical power of  
25 the laser light. The distribution of incident radiation doses at the sites of measurement can be precisely determined on the basis of the temporal distribution of the integral of fluorescence intensity.

We now describe an example of claim 23 with reference to Fig. 27. The purpose of this example is to measure radiation with the stimuable phosphor as a radiation detecting medium being provided at two or more sites. The radiation detecting portion comprises in superposition a plurality of stimuable phosphors, an optical fiber for illuminating each stimuable phosphor with exciting light, an optical bandpass filter centered at the wavelengths of prompt fluorescence and the stimulated fluorescence that is emitted from the stimuable phosphors upon illumination with exciting light, and at least one wavelength shifting fiber sensitive to the wavelength of fluorescence that is used to detect the emissions of the stimulated fluorescence and prompt fluorescence. To excite the respective stimuable phosphors, laser light is emitted from a multi-channel exciting light source and launched into the optical fibers. The sites of measurement can be selectively set by a control unit and stimulated fluorescence is read with a single wavelength shifting fiber. Signal processing is synchronous with the signal for controlling the exciting light source and performed by a conventional apparatus comprising a photomultiplier tube, a high-speed pulsed signal amplifier, a pulse height discriminator, a counter circuit and a control unit.

We now describe an example of claim 24 with reference to Fig. 28. A conventional apparatus for reading radiation image from a stimuable phosphor sheet (commercially available as an imaging plate) comprises a stimuable

phosphor sheet, an exciting light source emitting light of a wavelength that can excite the stimuable phosphor, a mechanism for illuminating the stimuable phosphor sheet with a rectangular pattern of the output exciting light, an  
5 optical bandpass filter centered at the wavelength of stimulated fluorescence, a wavelength shifter bundle comprising a ribbon array of wavelength shifting optical fibers that can be excited with the stimulated fluorescence, an optical bandpass filter centered at the  
10 wavelength of the shifted fluorescence, a photodetector capable of multi-channel detection of the fluorescence emitted from the respective wavelength shifting optical fibers, and a signal processing unit that processes the signals from the multi-channel detector to produce digital  
15 signals for constructing a radiation image. Claim 24 recites an improvement of this conventional apparatus. In order to illuminate the stimuable phosphor sheet with a rectangular pattern of the exciting light from the light source, laterally radiating optical fibers are arranged on  
20 the surface of the stimuable phosphor sheet in a direction perpendicular to the bundle of wavelength shifting optical fibers and the exciting light is launched from the light source into the laterally radiating optical fibers in turn. As a result, the mechanism for illuminating the exciting  
25 light is simplified and the dose of radiation accumulated in the stimuable phosphor sheet can be easily read together with the associated position information.

We now describe an example of claim 25 with reference

to Fig. 29. In this example, the fluorescence output from both ends of the wavelength shifting fiber in the radiation detecting portion described in claims 13 - 17 is passed through an optical bandpass filter centered at the  
5 wavelength of fluorescence and then detected with a photomultiplier tube used as the photodetector in the fluorescence detecting mechanism described in claim 10. By detecting the fluorescence output at both ends of the wavelength shifting fiber in the radiation detecting  
10 portion, the efficiency of detection is almost doubled.

We now describe an example of claim 26 with reference to Fig. 30. This example is the combination of claims 1 and 10 in that the method of radiation measurement recited in claim 1 is applied to the apparatus for radiation  
15 measurement recited in claim 10. A small radiation detecting portion that is capable of measurement over a long distance is operated by a method of measuring the incident radiation dose through selective detection of stimulated fluorescence and prompt fluorescence at  
20 specified time intervals and by a method capable of reading the dose of intense radiation incident within a very short time and, as a result, a high-performance apparatus for radiation measurement is realized.

We now describe an example of claim 27 with reference  
25 to Fig. 31. In this example, the signal reading method recited in claim 3 is substituted for the conventional pulse counting method in order to read the signal for stimulated fluorescence in the apparatus recited in

claim 26. The radiation detecting portion used in the example is a planar array comprising in superposition a bundle of laterally radiating optical fibers, a stimulable phosphor as a radiation detecting medium, an optical filter  
5 centered at the wavelength of fluorescence, and a bundle of wavelength shifting optical fibers sensitive to stimulated fluorescence. This radiation detecting portion is used to construct a differential and integral type of radiation measuring apparatus that selectively detects stimulated  
10 fluorescence and prompt fluorescence at specified time intervals on the basis of two actions of the stimulable phosphor, one for emitting stimulated fluorescence in proportion to the dose of incident radiation upon illumination with exciting light and the other for  
15 emitting prompt fluorescence in response to the incident radiation. The fluorescence detecting mechanism is monitored at specified time intervals with a saturation detecting circuit to see if it has failed by saturation with high-intensity radiation; if saturation occurs,  
20 stimulated fluorescence is detected after the detecting mechanism has recovered from saturation, thereby reading the dose of high-intensity radiation that was incident within a short time.

In this process, the quantity of exciting light to be  
25 applied to the stimulable phosphor is altered such that the dose of radiation can be measured without saturation of the fluorescence detecting mechanism. To read the dose of radiation, the stimulable phosphor is illuminated with

pulsed exciting light having a short time duration and the emission of stimulated fluorescence is detected with a photodetector, amplified by a charge-sensitive preamplifier, has the waveform shaped by a pulse shaping  
5 amplifier and fed into an analog/digital converter to determine the pulse height, on the basis of which a data collecting and processing circuit determines the dose of radiation accumulated in the stimuable phosphor.

We now describe an example of claim 28 with reference  
10 to Fig. 32. In this example, the stimuable phosphor used as a radiation detecting medium in the apparatus of claim 22 is replaced by a neutron imaging plate incorporating Gd which is a neutron converter capable of converting neutrons to an ionizable radiation (this plate is commercially  
15 available from Fuji Photo Film Co., Ltd. as one of BAS-ND Series neutron imaging plates). By using a radiation detecting medium which is a mixture of the neutron converter and the stimuable phosphor, one can construct a radiation measuring apparatus that performs the functions  
20 described in the previous examples and which has the added capability of measuring neutrons.

We now describe an example of claim 29 with reference to Fig. 33. In this example, a neutron imaging plate is used as the detection medium in the method of radiation  
25 measurement recited in claim 2 and a fast neutron moderator in the form of a warhead is provided in front of the imaging plate to allow for detection of fast neutrons. Exciting light is guided by an optical fiber and emitted

from within a space created by boring the center of the fast neutron moderator which is typically made of polyethylene.

We now describe an example of claim 30 with reference to Fig. 34. The emission characteristics of stimulated fluorescence from the stimuable phosphor as a radiation detecting medium and the phenomenon of fading (i.e., the radiation signal gradually disappears with time) both depend on temperature. To modify the radiation measuring apparatus according to the example of claim 26 on the basis of this finding, the temperature of the stimuable phosphor is measured with a temperature sensor and a temperature detecting circuit in the example of claim 30. On the basis of the measured temperature, the dose of accumulated radiation that is measured by illumination with exciting light is corrected by the data collecting and processing circuit to ensure precise measurement of radiation dose in spite of variations in the temperature of the radiation detecting portion.

#### 20 Effect of the Invention

Being constructed with these designs, the present invention offers the following advantages.

The dose of incident radiation is determined real-time from the results of measurement of stimulated fluorescence and prompt fluorescence at specified time intervals and this not only allows a wide dynamic range of radiations from a very weak level to a very intense level to be covered with a single detector but also enables

correct dosimetry of high-intensity radiations and neutrons that occur instantaneously.

The stimuable phosphor is illuminated with exciting light of short time duration and by taking advantage of the short lifetime of stimulated fluorescence, signal  
5 processing can be performed with a conventional radiation detecting system.

By using a laterally radiating optical fiber, the stimuable phosphor can be illuminated with exiting light  
10 by a simple and remote operation; there is no need to install a light source such as a laser in high-dose areas and one can construct a small mechanism for illuminating exciting light.

By a parallel arrangement of wavelength shifting  
15 fibers sensitive to the wavelengths of stimulated fluorescence and prompt fluorescence, a sufficient detection area is assured to enable the detection of stimulated fluorescence; in addition, there is no need to install a photodetector such as a photomultiplier tube in  
20 high-dose areas.

If desired, a ribbon array of wavelength shifting optical fibers may be used in combination with a parallel arrangement of laterally radiating optical fibers. By sandwiching the laterally radiating optical fibers between  
25 assemblies each consisting of a stimuable phosphor, an optical filter and the ribbon array of wavelength shifting optical fibers, small-sized radiation detecting portions can be constructed in various shapes for detection.



If desired, the detecting portion of the sandwich structure may be extended over a distance and exciting light having a very short pulse duration is launched into the laterally radiating optical fibers so that the incident  
5 exciting light propagates to illuminate the stimuable phosphor and the emission of stimulated fluorescence is detected with the bundle of wavelength shifting optical fibers. In this way, the positional distribution of radiation doses can be measured over a long distance.

10 The stimuable phosphor may incorporate, mix or combine with a neutron converter such as Gd that converts neutrons into an ionizable radiation. By using the thus prepared neutron detector, one can construct a radiation measuring apparatus capable of neutron detection.

15 By using the above-described apparatus and method for radiation measurement, a wide dynamic range of radiation doses from a very weak level to a very intense level can be covered with a single detector and dosimetry of high-intensity radiations and neutrons that occur  
20 instantaneously can be achieved; as a result, the doses of high-intensity radiations and neutrons that occur instantaneously around an accelerator or a target or those which result from an unexpected accident in nuclear reactor facilities can be measured conveniently, with high  
25 sensitivity and in high precision.